

SPECIFICATIONS FOR DI WATER SYSTEMS

All deionized water systems must meet the water quality standard specifications of the American Society for Testing Materials Type I and Type II water. Equipment must include five-micron prefilters, fresh activated carbon, strong base anion bed, strong acid/strong base mixed bed, and strong base Type I macroreticular anion exchange resin for colloidal removal.

Services must include 4-hour response time during normal working hours and call-back within 20 minutes for off hour response.

The vendor will perform initial tank set-up or routine tank replacement to include quality light and faucets in a manner that does not interfere with laboratory processes. Each system to include quality indicator light.

The number of filter units described is a maximum number.

The vendor will sanitize and install filter system for Type 1 water and pure lab systems within 24 hours of the request.

Immediate cancellation shall be administered when violations are found to be an impediment to the function of the agency and detrimental to its cause, or when conditions preclude the 30 day notice.

This contract shall be in force until expiration date or until 30 days after notice has been given by the state of Oklahoma of its desire to terminate the contract. After the first 6 months, vendor may cancel with 10 days written notice.

This contract is of an indefinite quantity and the state may, or may not, buy the quantity mentioned in this contract. Vendor must clear all shipments with agency prior to shipping any portion of this contract.

Request all items be awarded to one buyer.

Contract period 3/1/07-2/28/08

SPECIFICATIONS FOR DI WATER SYSTEMS

Item #	Quantity	Unit	Description	Unit Price	Extended Amount
1.	1	EA	INITIAL TANK SET-UP FOR THE DELIVERY OF ASTM TYPE II WATER AT 1 SINK. TANK SET-UP TO INCLUDE EACH: 0.45 CU FT MIXED BED RESIN TANK, 10" CARBON FILTER, PARTS, INSTALLATION MATERIAL AND LABOR. BRAND SIEMENS WATER TECHNOLOGY		
2.	65	TNK	0.45 CU FT MIXED BED RESIN TANK. BRAND: SWT # DIMB 10045FFP		
3.	80	EA	10" CARBON FILTER, BRAND: SWT # FCGCF 1020		
4.	15	EA	DISHWASHER FILTER AND TANK EXCHANGES TO INCLUDE: 9.75" 5 MICRON PRE-FILTER, BRAND: SWT # FCRON0905		
5.	10	TNK	1.2 CU FT CATION BED RESIN TANK, BRAND: SWT # DICATO120FSP		
6.	10	TNK	1.2 CU FT ANION BED RESIN TANK, BRAND: SWT # DIAN 10120FSP		
7.	15	TNK	1.2 CU FT MIXED BED RESIN TANK, BRAND: SWT # DIMB 10120 FSP		
8.	3	EA	PURELAB SYSTEM CARTRIDGE KITS, BRAND: SWT # PLC6000		
9.	2	EA	PURELAB UV BULB, BRAND: SWT # PL 2822		
10.	7	TNK	0.25 CU FT, CARBON TANK, BRAND: SWT # DICAR 0025 FSP		
11.	20	EA	0.79 CU FT, TYPE I (LOW TOC) ULTRA MIXED BED RESIN, BRAND: SWT # DIMB 10079 FSP		
12.	2	EA	METAL FREE GOOSE NECK FAUCETS CONNECTED TO SINKS		
13.	1	EA	PURELAB ULTRA ANALYTICAL ULTRAPURE POLISHING SYSTEM-REPLACEMENT: INCLUDES WALL BRACKET, INSTALLATION, LABOR AND FREIGHT.		
14.	2	EA	PURELAB ULTRATWIN PACK DI CARTRIDGE FOR SDI FEED, BRAND: SWT #LC148		
15.	5	EA	PURELAB SANITIZATION TABLETS. BRAND: SWT # CT1		
16.	15	EA	PURELAB FINAL FILTER. BRAND: SWT # LC 134		
17.	1	EA	PURELAB PLUS CPT BOARD, BRAND: SWT # PLP 231034		
18.	3	EA	PURELAB ULTRA - ULTRA FILTER BRAND: SWT # LC 109		
19.	3	EA	PURELAB ULTRA - UV BULB, BRAND: SWT # LC 118		
GRAND TOTAL					