

# PM-10 analyzers questions

PM-10 analyzers questions:

1. Will the analyzers be mounted in a rack or on a table top?
2. What length of Filterhead/Down tube? 1M, 1.5M, 2M?
3. Upon successful award of the equipment order when do you need delivery?

Answers:

1. Units will be tabletop mounted.
2. Would like 3 meters downtube – the downtube has to go thru the ceiling of our building and must be a minimum of 1 meter above the roof platform.
3. Delivery times are to be considered and evaluated with this bid. Please provide your best answer for the delivery date question when placing your bid.

Example:

